



# Certificate of Traceability

## Level I Laboratory Certification

This certifies that **Spectra Labs, Inc.**, using **Roche Cobas Integra 800 gen.2** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#7**

The system evaluated was:

Instrument: <b>Cobas Integra 800</b>	Calibrator Lot: <b>180338</b>	Application: <b>Gen.2</b>
Reagent Lot: <b>613420, 606212</b>	Calibrator Assigned Values: <b>2.704, 1.619, 1.352, 0.541, 0.108, 0.00 g/dL HbA1c; 2.6 g/dL, 12.7g/dL Hb</b>	

Date of Certification: August 1, 2009

Certification Expires: August 1, 2010

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor